

# DISSOLVED OXYGEN PROBE

Model : OXPB-09N, OXPB-11, OXPB-22

ISO-9001, CE, IEC1010



OXEL-03



OXHD-04

## Model : OXPB-09N

- \* Exclusive dissolved oxygen probe for Dissolved oxygen meter, DO-5509.
- \* Optional dissolved oxygen probe for Dissolved oxygen meter, DO-5509.
- \* Output is used 9 pins DIN plug.
- \* Cable : 4 meters.
- \* Optional accessories :  
DO probe filling electrolyte , OXEL-03.  
DO probe head diaphragm set , OXHD-04.

## Model : OXPB-11

- \* Exclusive dissolved oxygen probe for DO meter, DO-5510, DO-5519, DO-5510HA, PDO-520, DO-5512SD. YK-22DOA.
- \* Optional dissolved oxygen probe for DO meter, DO-5510, DO-5519, DO-5510HA, PDO-520, DO-5512SD. YK-22DOA.
- \* Output is used 4 pins DIN plug.
- \* Cable : 4 meters.
- \* Optional accessories :  
DO probe filling electrolyte , OXEL-03.  
DO probe head diaphragm set , OXHD-04.

## Model : OXPB-22

- \* Standard dissolved oxygen probe for Dissolved oxygen meter, YK-22DO.
- \* Optional dissolved oxygen probe for Dissolved oxygen meter, YK-22DO.
- \* Output is used the card plug.
- \* Cable : 4 meters.
- \* Optional accessories :  
DO probe filling electrolyte , OXEL-03.  
DO probe head diaphragm set , OXHD-04.

*@As intend to keep the DO probe under best condition, when user receive the probe ( OXPB-09N, OXPB-11, OXPB-22 ) , it should fill the Probe's Electrolyte ( OXEL-03 ) at first.*



LUTRON ELECTRONIC

**The Art of Measurement**